

भक्तपुर क्यान्सर अस्पताल
लागत अनुमानको लागि दर रेट पेश गर्ने बारे सूचना
(सूचना प्रकाशित मिति : २०८२/१२/२७)

यस अस्पतालको लागि यसै साथ संलग्न स्पेशिफिकेशन बमोजिमको Liquid Waste Management System खरीदको लागि लागत अनुमान तयार गर्न मिति २०८२/१२/३० सम्म यस अस्पतालको खरिद शाखामा दररेट पेश गर्नुहुन सूचित गरिन्छ ।

दररेट पेश गर्दा संलग्न गर्नुपर्ने कागजातहरू:

१. गत आ.व। २०८१/२०८२ सम्मको करचुक्ता प्रमाण पत्र
२. स्थायी लेखा नम्बर दर्ता प्रमाण पत्र
३. फर्म व्यवसाय दर्ता प्रमाणपत्र



Specification of Liquid Waste Treatment System

Device should perform decontamination and chemical neutralization to allow the safe discharge of all hazardous chemical and biological liquid medical wastes generated within the hospital into the municipal sewage system.

Minimum treatment capacity: more than 10,000 Liters per Day

The equipment should be of a Hazardous chemical addition free system.

The equipment should be an electrochemical technology for medical waste-water treatment without any chemical addition, simple and easy Operation with lower sludge production.

Equipment outer body surface to be made of Corrosion Resistant Materials grade body and tanks should be of corrosion proof material.

Must provide European CE certification for device (Machinery Directive).

Comprehensive Warranty for 5 years.

D. Stephan *[Signature]* *[Signature]* *[Signature]*
[Signature] *[Signature]*
[Signature]